

Ferrotech Verification Services
341 Nevada Crescent, Faerie Glen, Gauteng
082 650 0817

Page 1 of 1

VERIFICATION CERTIFICATE
Issued in terms of Regulation 120 of the legal metrology Regulations

Name of client Gauteng Metal Recyclers Serial no of this certificate 0021

Address 7 Lower Boksburg rd, Ext 10, Germiston

Number of laboratory verification procedure used: PROC 1

Description and identification of measuring instrument verified.

Make LMZ Model LS210 Capacity 40.000 Kg

Division size 20 Kg SA no 1531 210 AA 5 Serial no 11730

Description and serial no of components Tick applicable block Yes ☒ No ☐

Name component/s..... Type approval no/s.....

Serial no.....

Description of measurement standards utilized. ☒ Own or loan ☐ measurement standards

NOTE. The below mentioned verification standards are subject to an annual re-calibration process by a SANAS accredited calibration laboratory. The actual date of calibration is shown and also the date on which the calibration expires (12 months onwards from the date of calibration)

The measurement standards used are directly related to the reading or obtaining of the results of measurement and are traceable to the national standard.

<u>Serial no</u>	<u>Calibration cert no</u>	<u>Date of calibration of measurement standards</u>	<u>Expiry date</u>
<u>Description/set no</u>			
Small Set with 20 denominations.....	2019 - M 094.....	2019-10-07.....	2020-10-06.....
34 x 20kg + 1kg x 10kg.....	2019 - M 093.....	2019-10-03.....	2020-10-02.....
500kg Set 1 to 45.....	2019 - M 092.....	2019-10-03.....	2020-10-02.....

Verification Officer confirms that a visual inspection of the measurement standards was carried out prior to use to affirm fit for use. Tick ☒

Latest security intervention value – (where applicable) N/A.

Statement

The measuring instrument(s) was/were tested and found to comply in all respects with the requirements of the Legal Metrology Act, 2014 (Act No. 9 of 2014) and may be used for a prescribed purpose as intended by the Act.

Verification officer's initials and surname D.V Bezuidenhout Seal ID number of VO FER02

Verification officer's signature [Signature] Date of verification 18.08.2020

Serial no of verification mark FER02 Expiry date of this verification certificate 17.08.2022